



## **Specifications:**

**Stature:** Infant

**Range:**

Infant Circumference (05cm to 15cm)

Well nourished Infant Circumference (10cm to 30cm) when cascaded

**Bladder Width:** 5cm

**Bladder Volume:**

Restricted =

Non-Restricted=

**Material:**

Medical Grade Urethane

Nylon

Latex Free



## **Regulatory:**

AHA, AAMI Sp-10 and IEC-60601 compliant

FDA 510K

## **Coupling:**

Exclusive dual molded tubes form the female connector for use with TEC-Connect system or industry standard Rectus male connectors.

## **Minimum Service Life:**

Inflate 100,000 cycles @ 350mmHg in a restricted mode.

AAMI Sp-10 = (10,000 @ 330mmHg)

Inflate 1,000 cycles @ 250mmHg in a non-restrictive mode.

No AAMI specification

Functional hook and loop after 1,000 cycles @ 350mmHg.

AAMI Sp-10 = (1,000 @ no abuse pressure)

BP cuff transducer is fully functional after 60 washing cycles using automatic machines.